

? d s

Set Items Description

S1 1911725 S MRI OR MAGNETIC(1W)(IMAG? OR IMAGING) OR MAGNETIC(W)RESONAN? OR NMR OR
NUCLEAR()MAGNETIC()RESONANCE OR FTNMR OR FTMRI
S2 178391 S MAGNETORESONANCE OR PMR OR PROTON(W)MAGNETIC(W)RESONAN? OR MR()(IMAGE? OR
IMAGING)
S3 7517 S MC=(S01-E02A2 OR S03-E07A OR S01-E02A8A OR S01-E02A1 OR S03-E07C OR S05-D02B1 OR
S03-C02F1)
S4 49299 IC=(G01R-003 OR G01N-024/08 OR G01V-003/A75 OR G01R-033/56F OR G01V-003/00) FROM 2, 155,
5, 6, 8, 73, 94, 35, 144, 105, 99, 58, 34, 434, 292, 89, 65, 360, 239, 347, 305, 350, 162, 164, 357, 315, 23, 46, 68, 60, 33,
335, 294
S5 22261 S CC=(A0758 OR A8760I OR B7510N)
S6 1985792 S S1:S5
S7 13479 S MECHANICAL(2N)OSCILLAT?
S8 44 S STATE?(3N)VOXEL?
S9 441451 S NON()LINEAR?
S10 662434 S DISTORT?????
S11 2878 S ELASTO()GRAPH? OR ELASTOGRAPH?
S12 124 S S6 AND S7
S13 0 S S8 AND S12
S14 7 S S12 AND S11
S15 5 RD (unique items)
S16 122 S S9 AND S10 AND S6
S17 1 S S16 AND (S7 OR S8 OR S11)
S18 0 S S11 AND S7 AND S16
S19 0 S S11 AND S7 AND S9 AND S10
S20 10504 S S9 AND S10
S21 2 S S20 AND S11
S22 1 S S21 NOT S17
S23 622 S S11 AND S6
S24 8 S S23 AND (S7 OR S20)
S25 2 S S24 NOT (S15 OR S17)
S26 1 RD (unique items)
S27 0 S S23 AND VARIAT?(3N)TIM????
S28 9 S S23 AND S10
S29 5 RD (unique items)
S30 0 S S20 AND S12
S31 3788 S S9(2N)S10
S32 33 S S31 AND S6
S33 21 RD (unique items)
S34 20 S S33 NOT (S15 OR S17 OR S21 OR S22 OR S29)

? show files

[File 2] **INSPEC** 1969-2005/Sep W3

(c) 2005 Institution of Electrical Engineers. All rights reserved.

[File 155] **MEDLINE(R)** 1951-2005/Sep 27

(c) format only 2005 Dialog. All rights reserved.

[File 5] **Biosis Previews(R)** 1969-2005/Sep W4

(c) 2005 BIOSIS. All rights reserved.

[File 6] **NTIS** 1964-2005/Sep W3

(c) 2005 NTIS, Intl Cpyrght All Rights Res. All rights reserved.

[File 8] **Ei Compendex(R)** 1970-2005/Sep W3
(c) 2005 Elsevier Eng. Info. Inc. All rights reserved.

[File 73] **EMBASE** 1974-2005/Sep 28
(c) 2005 Elsevier Science B.V. All rights reserved.

[File 94] **JICST-EPlus** 1985-2005/Jul W5
(c)2005 Japan Science and Tech Corp(JST). All rights reserved.

[File 35] **Dissertation Abs Online** 1861-2005/Aug
(c) 2005 ProQuest Info&Learning. All rights reserved.

[File 144] **Pascal** 1973-2005/Sep W3
(c) 2005 INIST/CNRS. All rights reserved.

[File 105] **AESIS** 1851-2001/Jul
(c) 2001 Australian Mineral Foundation Inc. All rights reserved.
**File 105: This file is closed (no updates)*

[File 99] **Wilson Appl. Sci & Tech Abs** 1983-2005/Jul
(c) 2005 The HW Wilson Co. All rights reserved.

[File 58] **GeoArchive** 1974-2005/Jun
(c) 2005 Geosystems. All rights reserved.

[File 34] **SciSearch(R) Cited Ref Sci** 1990-2005/Sep W3
(c) 2005 Inst for Sci Info. All rights reserved.

[File 434] **SciSearch(R) Cited Ref Sci** 1974-1989/Dec
(c) 1998 Inst for Sci Info. All rights reserved.

[File 292] **GEOBASE(TM)** 1980-2005/Aug B1
(c) 2005 Elsevier Science Ltd. All rights reserved.

[File 89] **GeoRef** 1785-2005/Sep B1
(c) 2005 American Geological Institute. All rights reserved.
**File 89: Please see HELP ALERTALL for new Alert frequency and price. Please see HELP RATES 89 for new Academic Subscriber rates.*

[File 65] **Inside Conferences** 1993-2005/Sep W4
(c) 2005 BLDSC all rts. reserv. All rights reserved.

[File 360] **Specialty Chemicals Update Program** 2000/Q2
(c) 2000 SRI International. All rights reserved.
**File 360: Full fmts cost \$85.00 each for TYPEs, DISPLAYs, & PRINTs. Fmt 7 costs \$50.00. SCUP subscribers - use F960. Updating suspended.*

[File 239] **Mathsci** 1940-2005/Nov

(c) 2005 American Mathematical Society. All rights reserved.

[File 347] **JAPIO** Nov 1976-2005/Apr(Updated 050801)

(c) 2005 JPO & JAPIO. All rights reserved.

[File 305] **Analytical Abstracts** 1980-2005/Sep W2

(c) 2005 Royal Soc Chemistry. All rights reserved.

**File 305: Alert feature enhanced for multiple files, duplicate removal, customized scheduling. See HELP ALERT.*

[File 350] **Derwent WPIX** 1963-2005/UD,UM &UP=200561

(c) 2005 Thomson Derwent. All rights reserved.

**File 350: For more current information, include File 331 in your search. Enter HELP NEWS 331 for details.*

[File 162] **Global Health** 1983-2005/Aug

(c) 2005 CAB International. All rights reserved.

[File 164] **Allied & Complementary Medicine** 1984-2005/Sep

(c) 2005 BLHCIS. All rights reserved.

[File 357] **Derwent Biotech Res.** 1982-2005/Sep W4

(c) 2005 Thomson Derwent & ISI. All rights reserved.

[File 315] **ChemEng & Biotec Abs** 1970-2005/Aug

(c) 2005 DECHEMA. All rights reserved.

[File 23] **CSA Technology Research Database** 1963-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 46] **Corrosion Abstracts** 1966-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 68] **Solid State & Superconductivity Abstracts** 1966-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 60] **ANTE: Abstracts in New Tech & Engineer** 1966-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 33] **Aluminium Industry Abstracts** 1966-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 335] **Ceramic Abstracts/World Ceramics Abstracts** 1966-2005/Sep

(c) 2005 CSA. All rights reserved.

[File 294] **ONTAP(R) SciSearch(R) Cited Ref Science**

(c) 1991 Inst for Sci Info. All rights reserved.

17/9/1 (Item 1 from file: 350) Links
Derwent WPIX
(c) 2005 Thomson Derwent. All rights reserved.

015619236 **Image available**
WPI Acc No: 2003-681407/200365
XRPX Acc No: N03-544012

Magnetic resonance

**elastography for detection of different tissue types, e.g.
malignant lesions, whereby a patient or part of a patient is subject to
mechanical vibrations and the resulting non-
linear distortions evaluated**

Patent Assignee: PHILIPS INTELLECTUAL PROPERTY GMBH (PHIG); KONINK
PHILIPS ELECTRONICS NV (PHIG); SINKUS R (SINK-I); WEISS S (WEIS-I)
Inventor: SINKUS R; WEISS S
Number of Countries: 103 Number of Patents: 006
Patent Family:

Patent No	Kind	Date	Applicat No	Kind	Date	Week
DE 10209257	A1	20030911	DE 10209257	A	20020301	200365 B
WO 200373933	A1	20030912	WO 2003IB757	A	20030226	200370
AU 2003248594	A1	20030916	AU 2003248594	A	20030226	200430
EP 1482833	A1	20041208	EP 2003743465	A	20030226	200480
			WO 2003IB757	A	20030226	
US 20050104588	A1	20050519	WO 2003IB757	A	20030226	200534
			US 2004506456	A	20040831	
JP 2005518857	W	20050630	JP 2003572458	A	20030226	200543
			WO 2003IB757	A	20030226	

Priority Applications (No Type Date): DE 10209257 A 20020301

Patent Details:

Patent No	Kind	Lan	Pg	Main IPC	Filing Notes
-----------	------	-----	----	----------	--------------

DE 10209257	A1	10	G01N-003/32		
-------------	----	----	-------------	--	--

WO 200373933	A1	E	A61B-005/055		
--------------	----	---	--------------	--	--

Designated States (National): AE AG AL AM AT AU AZ BA BB BG BR BY BZ CA
CH CN CO CR CU CZ DE DK DM DZ EC EE ES FI GB GD GE GH GM HR HU ID IL IN
IS JP KE KG KP KR KZ LC LK LR LS LT LU LV MA MD MG MK MN MW MX MZ NO NZ
OM PH PL PT RO RU SC SD SE SG SK SL TJ TM TN TR TT TZ UA UG US UZ VC VN
YU ZA ZM ZW

Designated States (Regional): AT BE BG CH CY CZ DE DK EA EE ES FI FR GB
GH GM GR HU IE IT KE LS LU MC MW MZ NL OA PT SD SE SI SK SL SZ TR TZ UG
ZM ZW

AU 2003248594	A1		A61B-005/055	Based on patent WO 200373933
---------------	----	--	--------------	------------------------------

EP 1482833	A1	E	A61B-005/055	Based on patent WO 200373933
------------	----	---	--------------	------------------------------

Designated States (Regional): AL AT BE BG CH CY CZ DE DK EE ES FI FR GB
GR HU IE IT LI LT LU LV MC MK NL PT RO SE SI SK TR

US 20050104588	A1		G01V-003/00	
----------------	----	--	-------------	--

JP 2005518857	W	13	A61B-005/055	Based on patent WO 200373933
---------------	---	----	--------------	------------------------------

Abstract (Basic): DE 10209257 A1

NOVELTY - Object examination method has the following steps:
excitation of the object with mechanical vibrations; measurement of the
temporal behavior of object voxels that are displaced by waves in the

objects caused by the vibrations; determination of **non-linear distortion** from the temporal behavior of the deflections; and evaluation of the **non-linear distortions**.

DETAILED DESCRIPTION - INDEPENDENT CLAIMS are included for:

(1) arrangement for implementation of the method;

(2) a computer program for use with a **magnetic resonance** instrument to implement the method.

USE - **Magnetic resonance**

elastography for detection of different tissue types, e.g. malignant lesions.

ADVANTAGE - The inventive method yields additional information that enables better differentiation between tissue types.

DESCRIPTION OF DRAWING(S) - Figure shows a sequence of vibrations for use with MR data acquisition.

pp; 10 DwgNo 4/6

Title Terms: MAGNETIC; RESONANCE; DETECT; TISSUE; TYPE; MALIGNANT; LESION; PATIENT; PART; PATIENT; SUBJECT; MECHANICAL; VIBRATION; RESULT; NON; LINEAR; **DISTORT**; EVALUATE

Derwent Class: P31; S01; S03; S05; T01

International Patent Class (Main): A61B-005/055; G01N-003/32;

G01V-003/00

International Patent Class (Additional): A61B-008/08; G01M-007/00;

G01R-033/28

File Segment: EPI; EngPI

Manual Codes (EPI/S-X): **S01-E02A2**; **S03-E07A**;

S05-D01J; S05-D02B2; T01-J06A; T01-S03

22/9/1 (Item 1 from file: 34) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

SciSearch(R) Cited Ref Sci

(c) 2005 Inst for Sci Info. All rights reserved.

10385045 **Genuine Article#:** 521LE **Number of References:** 38

Advances in ultrasound

Author: Claudon M (REPRINT) ; Tranquart F; Evans DH; Lefevre F; Correas JM

Corporate Source: Univ Nancy,F-54511 Vandoeuvre Les Nancy//France/ (REPRINT); Univ Nancy,F-54511

Vandoeuvre Les Nancy//France/; Univ Tours,F-37044 Tours//France/; Univ Leicester,Leicester LE1

5WW/Leics/England/; Univ Paris Necker,F-75730 Paris//France/; Hop Enfants,Serv Radiol,F-54511 Vandoeuvre

Les Nancy//France/

Journal: EUROPEAN RADIOLOGY , 2002 , V 12 , N1 (JAN) , P 7-18

ISSN: 0938-7994 **Publication date:** 20020100

Publisher: SPRINGER-VERLAG , 175 FIFTH AVE, NEW YORK, NY 10010 USA

Language: English **Document Type:** ARTICLE

Geographic Location: France; England

Journal Subject Category: RADIOLOGY, NUCLEAR MEDICINE & MEDICAL IMAGING

Abstract: Significant advances have been recently introduced into various fields of technology, taking advantage of the use of new piezoelectric materials and the large diffusion of broadband transducers. Various types of modulation may be applied to the pulse characteristics, using single pulse, multipulse or multiline techniques, and resulting in improved spatial resolution and better penetration. **Non-linear** imaging uses the harmonics component, which is generated by tissues or by contrast agents. Different modalities can be used to separate harmonics from fundamental bands from the received signal. New Doppler modes have been developed, whereas grey-scale flow imaging allows the simultaneous imaging of blood flow and tissues. Compounding techniques improve the contrast resolution of tissues and reduce artefacts. If 3D techniques are now currently available, real-time 4D imaging has been recently introduced. **Elastographic** imaging is still under evaluation, but promising clinical results have been shown. Recent release of the DICOM specification has made the full integration of ultrasound to the PALS systems easier. All these advances indicate that the contribution and potential of ultrasound in patient management is still growing.

Descriptors--Author Keywords: US ; technology ; Doppler modes

Identifiers-- Key Word Plus(R): BLOOD-FLOW; B-FLOW; NONLINEARITY; PROPAGATION; TRANSDUCERS; **DISTORTION**; PROSTATE

Cited References:

BAMBER JC, 1996, V22, P285, ACOUST IMAG

BEGA G, 2000, V22, P13, DIAGN IMAGING S

BERSON M, 1999, V10, P53, EUR J ULTRASOUND

BRUINING N, 2000, V16, P31, DIAGN IMAGING EUR

BURNS PN, 1996, V51, P50, CLIN RADIOL S1

CHIAO R, 2000, IEEE INT ULTR S PUER

CHRISTOPHER T, 1997, V44, P125, IEEE T ULTRASON FERR

CORREAS JM, 2001, V11, P1316, EUR RADIOL

DAWSON P, 1999, P451, TXB CONTRAST MEDIA

DEKORTE CL, 1997, V23, P735, ULTRASOUND MED BIOL

FILIPCZYNSKI L, 1999, V25, P285, ULTRASOUND MED BIOL

FOSTER FS, 2000, V26, P1, ULTRASOUND MED BIOL

FREIHERR G, 1998, PHU6, DIAGNOSTIC IMAGI JUN

HAMPER UM, 1999, V212, P719, RADIOLOGY

HENRI P, 2000, V81, P465, J RADIOL
KALLEL F, 1999, V21, P201, ULTRASONIC IMAGING
KORNBLUTH M, 1998, V11, P693, J AM SOC ECHOCARDIOG
LAW WK, 1985, V11, P307, ULTRASOUND MED BIOL
LIU JB, 1999, V25, P159, ULTRASOUND MED BIOL
LOWERS J, 2000, V22, P63, DIAGN IMAGING
MILLER DL, 1998, V104, P2498, J ACOUST SOC AM
MUIR TG, 1997, V6, P345, ULTRASOUND MED BIOL
NELSON TR, 1998, V24, P1243, ULTRASOUND MED BIOL
NICHOLS S, 2000, V22, P65, DIAGN IMAGING
OPHIR J, 1999, V213, P203, P I MECH ENG H
RIZZATTO G, 1999, V9, PS304, EUR RADIOL S3
ROSENTHAL SJ, 2001, V176, P1393, AM J ROENTGENOL
RUBIN DN, 1998, V32, PA127, J AM COLL CARDIOL
SCHROPE B, 1992, V14, P134, ULTRASONIC IMAGING
SEHGAL CM, 1986, V12, P865, ULTRASOUND MED BIOL
STARRITT HC, 1991, V36, P1485, PHYS MED BIOL
STARRITT HC, 1986, V31, P1401, PHYS MED BIOL
TRANQUART F, 1999, V25, P889, ULTRASOUND MED BIOL
WARD B, 1997, V101, P143, J ACOUST SOC AM
WESKOTT HP, 2000, V21, P59, ULTRASCHALL MED
WHITTINGHAM TA, 1999, V9, PS298, EUR RADIOL S3
WHITTINGHAM TA, 1999, V9, PS307, EUR RADIOL S3
WHITTINGHAM TA, 1999, V9, PS323, EUR RADIOL S3

26/9/1 (Item 1 from file: 73) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

EMBASE

(c) 2005 Elsevier Science B.V. All rights reserved.

12769467 EMBASE No: 2004360821

MR elastography of the prostate: Initial in-vivo application

Kemper J.; Sinkus R.; Lorenzen J.; Nolte-Ernsting C.; Stork A.; Adam G.

Dr. J. Kemper, Klin. Diagn. und Interventionelle R., Universitätsklinikum Eppendorf (UKE), Martinistr. 52, 20246 Hamburg Germany

Author Email: jkemper@uke.uni-hamburg.de

RoFo Fortschritte auf dem Gebiet der Röntgenstrahlen und der Bildgebenden Verfahren (ROFO FORTSCHR. GEB. RONTGENSTR. BILDGEBENDEN VERFAHREN) (Germany) 2004 , 176/8 (1094-1099)

CODEN: RFGVE **ISSN:** 1438-9029

Document Type: Journal ; Article

Language: ENGLISH **Summary Language:** ENGLISH; GERMAN

Number Of References: 20

Purpose: To analyze the initial assessment of the technical feasibility of in-vivo MR **elastography** (MRE) of the prostate gland in healthy volunteers. **Materials and Methods:** Dynamic sinusoidal MR **elastography** was performed in 7 healthy volunteers in prone position. The mechanical wave was induced via an external oscillator attached to the pubic bone. A 1.5 Tesla MR system (Philips Medical Systems, Netherland) was used with 4 combined surface coils for signal reception. MRE data acquisition was performed with a motion-sensitive spin-echo MR sequence that was phase-locked to the **mechanical oscillation**. Subsequently, these images were used to reconstruct the local distribution of elasticity inside the prostate gland. The applied reconstruction algorithm was tested by means of phantom measurements. **Results:** Sufficient penetration of the mechanical wave into the prostate gland was achieved in all volunteers, allowing the acquisition of utilizable image data sets. The reconstructed distribution of elasticity (shear-modulus) inside the healthy prostate gland correlated with the zonal anatomy of the gland. The elasticity of the central portion (2.2 +/- 0.3 kPa) appeared to be lower than the peripheral prostatic portion (3.3 +/- 0.5 kPa). **Conclusion:** In-vivo MRE of the prostate gland is technically feasible. The proposed experimental set-up allows the efficient insertion of the mechanical wave into the prostate gland and provides a successful MR data acquisition.

Device Brand Name/Manufacturer Name: 1.5 Tesla MR system/Philips/Netherlands

Device Manufacturer Names: Philips/Netherlands

MEDICAL DESCRIPTORS:

*** elastography; *nuclear magnetic resonance imaging ; *prostatefeasibility stud**

feasibility study; oscillator; pubic bone; oscillation; image reconstruction; elasticity; in vivo study; human; male; human experiment; normal human; adult; article; priority journal

SECTION HEADINGS:

014 Radiology

028 Urology and Nephrology

29/9/1 (Item 1 from file: 2) [Links](#)

Fulltext available through: [Institute of Electrical and Electronics Engineers](#) [USPTO Full Text Retrieval Options](#)
INSPEC

(c) 2005 Institution of Electrical Engineers. All rights reserved.

08674628 INSPEC Abstract Number: A2003-16-8760I-012, B2003-08-7510N-047

Title: Measurement of in vivo local shear modulus using MR elastography multiple-phase patchwork offsets

Author Suga, M.; Matsuda, T.; Minato, K.; Oshiro, O.; Chihara, K.; Okamoto, J.; Takizawa, O.; Komori, M.; Takahashi, T.

Author Affiliation: Graduate Sch. of Inf. Sci., Nara Inst. of Sci. & Technol., Japan

Journal: IEEE Transactions on Biomedical Engineering vol.50, no.7 p. 908-15

Publisher: IEEE ,

Publication Date: July 2003 **Country of Publication:** USA

CODEN: IEBEAX **ISSN:** 0018-9294

SICI: 0018-9294(200307)50:7L:908:MVL5;1-I

Material Identity Number: I050-2003-007

U.S. Copyright Clearance Center Code: 0018-9294/03/\$17.00

Language: English **Document Type:** Journal Paper (JP)

Treatment: Theoretical (T)

Abstract: Magnetic resonance elastography (MRE) is a method that can visualize the propagating and standing shear waves in an object being measured. The quantitative value of a shear modulus can be calculated by estimating the local shear wavelength. Low-frequency mechanical motion must be used for soft, tissue-like objects because a propagating shear wave rapidly attenuates at a higher frequency. Moreover, a propagating shear wave is **distorted** by reflections from the boundaries of objects. However, the **distortions** are minimal around the wave front of the propagating shear wave. Therefore, we can avoid the effect of reflection on a region of interest (ROI) by adjusting the duration of mechanical vibrations. Thus, the ROI is often shorter than the propagating shear wavelength. In the MRE sequence, a motion-sensitizing gradient (MSG) is synchronized with mechanical cyclic motion. MRE images with multiple initial phase offsets can be generated with increasing delays between the MSG and mechanical vibrations. This paper proposes a method for measuring the local shear wavelength using MRE multiple initial phase patchwork offsets that can be used when the size of the object being measured is shorter than the local wavelength. To confirm the reliability of the proposed method, computer simulations, a simulated tissue study and in vitro and in vivo studies were performed. (17 Refs)

Descriptors: biomechanics; biomedical MRI; elastic moduli measurement; shear modulus

Identifiers: in vivo local shear modulus measurement; object boundaries; local wavelength; region of interest; mechanical vibrations duration; **magnetic resonance imaging**; medical diagnostic imaging; motion-sensitizing gradient; multiple initial phase patchwork offsets; in vivo human tissue; computer simulations; simulated tissue study; in vivo studies; in vitro studies

Class Codes: A8760I (Medical magnetic resonance imaging and spectroscopy); A8745B (Mechanical properties of tissues and organs); A8770E (Patient diagnostic methods and instrumentation); A8740 (Biomagnetism); B7510N (Biomedical magnetic resonance imaging and spectroscopy); B7320G (Mechanical variables measurement

Copyright 2003, IEE

29/9/2 (Item 1 from file: 8) [Links](#)

Ei Compendex(R)

(c) 2005 Elsevier Eng. Info. Inc. All rights reserved.

06930618 E.I. No: EIP04288254794

Title: Quantification of soft tissue deformations using strain-encoded magnetic resonance imaging

Author: Fahmy, Ahmed S.; Osman, Nael F.

Corporate Source: Electrical and Computer Engineering Johns Hopkins University, Baltimore, MD, United States

Conference Title: Medical Imaging 2004: Physiology, Function, and Structure from Medical Images

Conference Location: San Diego, CA, United States **Conference Date:** 20040215-20040217

Sponsor: SPIE

E.I. Conference No.: 63197

Source: Proceedings of SPIE - The International Society for Optical Engineering Medical Imaging 2004: Physiology, Function, and Structure from Medical Images v 5369 2004.

Publication Year: 2004

CODEN: PSISDG **ISSN:** 0277-786X

Language: English

Document Type: CA; (Conference Article) **Treatment:** T; (Theoretical)

Journal Announcement: 0407W2

Abstract: Strain Encoded **Magnetic Resonance Imaging** (SENC-MRI) is a new technique that allows real-time quantification of tissue deformation. The technique is based on initially modulating the magnetization of the imaged object with sinusoidal pattern (MR-tagging) in the z-direction (throughplane direction). Compression is then applied to the object resulting in a change of the frequency of the sinusoidal tagging depending, in part, on tissue stiffness; e.g. the softer the material the higher the resulting frequency. By determining the changes in frequency, regional deformations can be determined and quantified. In SENC MRI, this is achieved by acquiring several images (typically 8 images), each with different phase-encoding, which we call tunings, in the z-direction. For each tuning, the intensity of pixels whose tagging frequency coincides with the tuning frequency is higher than other pixels. Since the number of the acquired images is limited, only a limited range of frequencies can be covered and, hence, the accuracy of the estimates may be inefficient if the tuning are not selected carefully. However, in this paper, we show that deformation maps can be obtained with good accuracy from the limited number of tunings. In this regard, we propose three methods and compare between them for maximum achievable accuracy. The methods are 1) center-of-mass, 2.) curve fitting, and 3) clustering-based method. The methods are applied to simulated data and MR images obtained from a gel phantom experiment. The results of comparisons shows that good estimates of deformation can be obtained even if the sampled data is **distorted** by noise or MR artifacts. 15 Refs.

Descriptors: *Magnetic resonance imaging; Tissue; Tumors; Elastic moduli; Magnetization; Curve fitting; Sensitivity analysis; Parameter estimation; Mathematical models

Identifiers: Strain encoded imaging; Tissue deformation; Tissue stiffness; **Elastography**

Classification Codes:

701.2 (Magnetism, Basic Concepts & Phenomena); 461.2 (Biological Materials); 921.6 (Numerical Methods); 731.1 (Control Systems)

701 (Electricity & Magnetism); 461 (Bioengineering); 921 (Applied Mathematics); 731 (Automatic Control Principles & Applications)

70 (ELECTRICAL ENGINEERING, GENERAL); 46 (BIOENGINEERING); 92 (ENGINEERING MATHEMATICS); 73 (CONTROL ENGINEERING)

29/9/3 (Item 2 from file: 8) [Links](#)

Fulltext available through: [Institute of Electrical and Electronics Engineers](#) [USPTO Full Text Retrieval Options](#)
Ei Compendex(R)

(c) 2005 Elsevier Eng. Info. Inc. All rights reserved.

06873611 E.I. No: EIP04228189306

Title: Supersonic shear imaging: A new technique for soft tissue elasticity mapping

Author: Bercoff, Jeremy; Tanter, Mickael; Fink, Mathias

Corporate Source: Laboratoire Ondes et Acoustique Ecl. Sup. de Phys./de Chim. Indust. Universite Paris VII, 75005, Paris, France

Source: IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control v 51 n 4 April 2004. p 396-409

Publication Year: 2004

CODEN: ITUCER **ISSN:** 0885-3010

Language: English

Document Type: JA; (Journal Article) **Treatment:** T; (Theoretical)

Journal Announcement: 0406W2

Abstract: Supersonic shear imaging (SSI) is a new ultrasound-based technique for real-time visualization of soft tissue viscoelastic properties. Using ultrasonic focused beams, it is possible to remotely generate mechanical vibration sources radiating low-frequency, shear waves inside tissues. Relying on this concept, SSI proposes to create such a source and make it move at a supersonic speed. In analogy with the "sonic boom" created by a supersonic aircraft, the resulting shear waves will interfere constructively along a Mach cone, creating two intense plane shear waves. These waves propagate through the medium and are progressively **distorted** by tissue heterogeneities. An ultrafast scanner prototype is able to both generate this supersonic source and image (5000 frames/s) the propagation of the resulting shear waves. Using inversion algorithms, the shear elasticity of medium can be mapped quantitatively from this propagation movie. The SSI enables tissue elasticity mapping in less than 20 ms, even in strongly viscous medium like breast. Modalities such as shear compounding are implementable by tilting shear waves in different directions and improving the elasticity estimation. Results validating SSI in heterogeneous phantoms are presented. The first in vivo investigations made on healthy volunteers emphasize the potential clinical applicability of SSI for breast cancer detection. 16 Refs.

Descriptors: *Imaging systems; Tissue; Elasticity; Shear waves; **Magnetic resonance imaging**; Vibrations (mechanical); Acoustics; Algorithms; Mathematical models

Identifiers: Supersonic shear imaging; Soft tissue elasticity mapping; **Elastography**; Breast cancer

Classification Codes:

723.2 (Data Processing); 461.2 (Biological Materials); 931.1 (Mechanics) ; 701.2 (Magnetism, Basic Concepts & Phenomena)

723 (Computer Software, Data Handling & Applications); 461 (Bioengineering); 421 (Strength of Building Materials; Mechanical Properties); 931 (Applied Physics Generally); 701 (Electricity & Magnetism); 751 (Acoustics, Noise & Sound); 921 (Applied Mathematics)

72 (COMPUTERS & DATA PROCESSING); 46 (BIOENGINEERING); 42 (BUILDING MATERIALS PROPERTIES & TESTING); 93 (ENGINEERING PHYSICS); 70 (ELECTRICAL ENGINEERING, GENERAL); 75 (SOUND & ACOUSTICAL TECHNOLOGY); 92 (ENGINEERING MATHEMATICS)

29/9/4 (Item 3 from file: 8) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

Ei Compendex(R)

(c) 2005 Elsevier Eng. Info. Inc. All rights reserved.

05654705 E.I. No: EIP00095338205

Title: Tissue mechanical attributes imaging: principles and methods

Author: Kallel, Faouzi; Ophir, Jonathan

Corporate Source: Univ of Texas Medical Sch, Houston, TX, USA

Conference Title: CBMS 2000: 13th IEEE Symposium on Computer-Based Medical Systems

Conference Location: Houston, TX, USA **Conference Date:** 19000622-19000624

Sponsor: IEEE

E.I. Conference No.: 57265

Source: Proceedings of the IEEE Symposium on Computer-Based Medical Systems 2000. IEEE, Los Alamitos, CA, USA. p 147-159

Publication Year: 2000

CODEN: PSCSFM **ISSN:** 1063-7125

Language: English

Document Type: CA; (Conference Article) **Treatment:** T; (Theoretical)

Journal Announcement: 0010W4

Abstract: We review the principles of MR and ultrasonic approaches proposed for imaging the mechanical attributes or properties of tissues. We identify two groups of methods. One group of methods is based on the use of applied vibration to obtain information about tissue mechanical properties. The second group of methods is based on the application of quasi-static compression as a stimulus to map distribution of tissue mechanical attributes. Appropriate theoretical models are presented to explain both groups of methods. Finally, the principles of **elastography** are detailed with emphasis on its potential clinical applications. (Author abstract) 37 Refs.

Descriptors: *Medical imaging; Tissue; **Magnetic resonance imaging**; Ultrasonic imaging; Mathematical models; Acoustic waves; Signal **distortion**; Ultrasonic refraction

Identifiers: Tissue mechanical attributes imaging; Applied vibration; Quasi states compression; Stimulus; Map distribution; **Elastography**; Clinical applications

Classification Codes:

461.1 (Biomedical Engineering); 461.2 (Biological Materials); 701.2 (Magnetism: Basic Concepts & Phenomena); 753.3 (Ultrasonic Applications); 921.6 (Numerical Methods); 751.1 (Acoustic Waves)

461 (Biotechnology); 701 (Electricity & Magnetism); 753 (Sound Technology & Ultrasonics); 921 (Applied Mathematics); 751 (Acoustics)

46 (BIOENGINEERING); 70 (ELECTRICAL ENGINEERING); 75 (ACOUSTICAL TECHNOLOGY); 92 (ENGINEERING MATHEMATICS)

34/9/1 (Item 1 from file: 2) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

INSPEC

(c) 2005 Institution of Electrical Engineers. All rights reserved.

08178324 **INSPEC Abstract Number:** A2002-06-0758-006, B2002-03-7510N-047, C2002-03-5260B-441

Title: Non-linear compensation of distortions introduced by the presence of metal objects in magnetic resonance imaging

Author Bui, F.M.; Bott, K.; Mintchev, M.P.

Author Affiliation: Dept. of Electr. & Comput. Eng., Calgary Univ., Alta., Canada

Journal: Proceedings of the SPIE - The International Society for Optical Engineering **Conference Title:** Proc. SPIE - Int. Soc. Opt. Eng. (USA) vol.4322, pt.1-3 p. 845-52

Publisher: SPIE-Int. Soc. Opt. Eng ,

Publication Date: 2001 **Country of Publication:** USA

CODEN: PSISDG **ISSN:** 0277-786X

SICI: 0277-786X(2001)4322:1/3L.845:LCDI;1-4

Material Identity Number: C574-2001-256

U.S. Copyright Clearance Center Code: 0277-786X/01/\$15.00

Conference Title: Medical Imaging 2001: Image Processing

Conference Sponsor: SPIE

Conference Date: 19-22 Feb. 2001 **Conference Location:** San Diego, CA, USA

Language: English **Document Type:** Conference Paper (PA); Journal Paper (JP)

Treatment: Practical (P)

Abstract: Magnetic resonance imaging (MRI) exhibits technical characteristics that make it invaluable in medical diagnosis. However, its full potential has been severely limited by the presence of imaging artifacts. These artifacts cause distortions in the obtained images, which no longer faithfully represent the object being imaged. In this work, we study the particular type of artifact known as "magnetic susceptibility difference artifacts", caused by the presence of a ferromagnetic source. Previously (2000), we have quantified these artifacts from three different perspectives: (1) pixel displacement, (2) blurring and (3) nonlinearity. However, the nonlinear distortions were quantified using a single-parameter nonlinear function. In this study, a multi-parameter nonlinear function based on polynomial series is used in the optimization. The results show that this modification allows for more flexibility and, as a result, improves the optimization algorithm. (25 Refs)

Descriptors: error compensation; ferromagnetic materials; image processing; **magnetic resonance imaging**; magnetic susceptibility; metals; nonlinear distortion; optimisation; series (mathematics)

Identifiers: nonlinear distortion compensation; metal objects; **magnetic resonance imaging**; **MRI**; medical diagnosis; imaging artifacts; magnetic susceptibility difference artifacts; ferromagnetic source; pixel displacement; blurring; multi-parameter nonlinear function; polynomial series; optimization algorithm; flexibility; digital image processing; Taylor's series expansion; nonlinear signal processing

Class Codes: **A0758** (Magnetic resonance spectrometers, auxiliary instruments and techniques); **A7530C** (Magnetic moments and susceptibility in magnetically ordered materials); **A4230V** (Image processing and restoration); **A8760I** (Medical magnetic resonance imaging and spectroscopy); **B7510N** (Biomedical magnetic resonance imaging and spectroscopy); **B3110C** (Ferromagnetic materials); **B0260** (Optimisation techniques); **B6135** (Optical, image and video signal processing); **C5260B** (Computer vision and image processing techniques); **C1180** (Optimisation techniques); **C7330** (Biology and medical computing)

Copyright 2002, IEE

34/9/2 (Item 2 from file: 2) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

INSPEC

(c) 2005 Institution of Electrical Engineers. All rights reserved.

08000118 **INSPEC Abstract Number:** A2001-18-8760I-001, B2001-09-7510N-076, C2001-09-7330-286

Title: Volterra series modelling and compensation of non- linear distortions caused by susceptibility difference artefacts related to the presence of ferromagnetic implants in magnetic resonance imaging

Author Bui, F.M.; Jianchuan Li; Bott, K.; Mintchev, M.P.

Author Affiliation: Dept. of Electr. Eng., Calgary Univ., Alta., Canada

Journal: Medical Engineering & Physics vol.23, no.3 p. 207-15

Publisher: Elsevier ,

Publication Date: April 2001 **Country of Publication:** UK

CODEN: MEPHEO **ISSN:** 1350-4533

SICI: 1350-4533(200104)23:3L:207:VSMC;1-N

Material Identity Number: B478-2001-007

U.S. Copyright Clearance Center Code: 1350-4533/2001/\$20.00

Document Number: S1350-4533(01)00044-3

Language: English **Document Type:** Journal Paper (JP)

Treatment: Theoretical (T); Experimental (X)

Abstract: **Magnetic resonance imaging** is popular in medical settings due to its unique technical characteristics. However, its full potential has been limited by imaging artefacts caused by various phenomena. Previously, a methodology was proposed to characterize and reduce artefacts caused specifically by magnetic susceptibility differences. In the present work, the Volterra series approach is suggested as an alternative method for describing **non- linear distortions** induced by susceptibility artefacts. A second-order Volterra series is utilized for characterizing the image non-linearities using a block-by-block processing approach. Subsequently, a corresponding second-order inverse Volterra series is applied to compensate for the quantified distortions. In addition, a technique for automatic demarcation of recoverable and non-recoverable regions in **magnetic resonance images** is proposed to ameliorate the developed methodology. Experimental results indicate that this approach offers improved accuracy and flexibility in reducing nonlinear distortions caused by magnetic susceptibility artefacts. (27 Refs)

Descriptors: biomedical **MRI**; magnetic susceptibility; medical image processing; modelling; prosthetics; Volterra series

Identifiers: Volterra series modelling; nonlinear distortions compensation; ferromagnetic implants; medical **magnetic resonance imaging**; nonlinear distortions; block-by-block processing approach; medical diagnostic imaging; second-order inverse Volterra series; nonrecoverable regions; recoverable regions; nonlinear system identification; digital image processing

Class Codes: **A8760I** (Medical magnetic resonance imaging and spectroscopy); **A8740** (Biomagnetism); **A8770J** (Prosthetics and other practical applications); **A8770E** (Patient diagnostic methods and instrumentation); **A8710** (General, theoretical, and mathematical biophysics); **B7510N** (Biomedical magnetic resonance imaging and spectroscopy); **B7520E** (Prosthetics and orthotics); **B6135** (Optical, image and video signal processing); **C7330** (Biology and medical computing); **C5260B** (Computer vision and image processing techniques

Copyright 2001, IEE

34/9/3 (Item 3 from file: 2) **Links**

INSPEC

(c) 2005 Institution of Electrical Engineers. All rights reserved.

07630850 **INSPEC Abstract Number:** A2000-15-8760I-050, B2000-08-7510N-072, C2000-08-7330-163

Title: Multiresolution spline warping for EPI registration

Author Kybic, J.; Thevenaz, P.; Unser, M.

Author Affiliation: Biomed. Imaging Group, Swiss Fed. Inst. of Technol., Lausanne, Switzerland

Journal: Proceedings of the SPIE - The International Society for Optical Engineering **Conference Title:** Proc. SPIE - Int. Soc. Opt. Eng. (USA) vol.3813 p. 571-9

Publisher: SPIE-Int. Soc. Opt. Eng ,

Publication Date: 1999 **Country of Publication:** USA

CODEN: PSISDG **ISSN:** 0277-786X

SICI: 0277-786X(1999)3813L:571:MSWR;1-O

Material Identity Number: C574-1999-324

U.S. Copyright Clearance Center Code: 0277-786X/99/\$10.00

Conference Title: Wavelet Applications in Signal and Image Processing VII

Conference Sponsor: SPIE

Conference Date: 19-23 July 1999 **Conference Location:** Denver, CO, USA

Language: English **Document Type:** Conference Paper (PA); Journal Paper (JP)

Treatment: Theoretical (T)

Abstract: Registration of images subject to non-linear warping has numerous practical applications. We present an algorithm based on double multiresolution structure of warp and image spaces. Tuning a so-called scale parameter controls the coarseness of the grid by which the deformation is described and also the amount of implicit regularization. The application of our algorithm deals with undoing unidirectional **non- linear** geometrical **distortion** of echo-planar images (EPI) caused by local magnetic field inhomogeneities induced mainly by the subject presence. The unwarping is based on registering the EPI images with corresponding undistorted anatomical **MRI** images. We present evaluation of our method using a wavelet-based random Sobolev-type deformation generator as well as other experimental examples. (25 Refs)

Descriptors: biomedical **MRI**; brain; image registration; image resolution; medical image processing; splines (mathematics)

Identifiers: multiresolution spline warping; EPI registration; nonlinear warping; double multiresolution structure; image spaces; scale parameter; grid coarseness; regularization; unidirectional nonlinear geometrical distortion; echo-planar images; local magnetic field inhomogeneities; subject presence; wavelet-based random Sobolev-type deformation generator

Class Codes: **A8760I** (Medical magnetic resonance imaging and spectroscopy); **A8770E** (Patient diagnostic methods and instrumentation); **A8740** (Biomagnetism); **B7510N** (Biomedical magnetic resonance imaging and spectroscopy); **B6135** (Optical, image and video signal processing); **B0290F** (Interpolation and function approximation (numerical analysis)); **C7330** (Biology and medical computing); **C5260B** (Computer vision and image processing techniques); **C4130** (Interpolation and function approximation (numerical analysis)

Copyright 2000, IEE

34/9/4 (Item 4 from file: 2) [Links](#)

INSPEC

(c) 2005 Institution of Electrical Engineers. All rights reserved.

07497845 **INSPEC Abstract Number:** A2000-06-8760I-007, B2000-03-7510N-082, C2000-03-7330-300

Title: Method for partial reduction of the non-linearity and distortion in MR images caused by the presence of metal objects

Author Domoustchiev, K.; Bott, K.; Smith, M.R.; Mintchev, M.P.

Author Affiliation: Dept. of Electr. & Comput. Eng., Calgary Univ., Alta., Canada

Conference Title: Engineering Solutions for the Next Millennium. 1999 IEEE Canadian Conference on Electrical and Computer Engineering (Cat. No.99TH8411) **Part** vol.2 **p.** 851-4 vol.2

Editor(s): Meng, M.

Publisher: IEEE , Piscataway, NJ, USA

Publication Date: 1999 **Country of Publication:** USA 3 vol. (xxiii+1758) pp.

ISBN: 0 7803 5579 2 **Material Identity Number:** XX-1999-02873

U.S. Copyright Clearance Center Code: 0 7803 5579 2/99/\$10.00

Conference Title: Engineering Solutions for the Next Millennium. 1999 IEEE Canadian Conference on Electrical and Computer Engineering

Conference Date: 9-12 May 1999 **Conference Location:** Edmonton, Alta., Canada

Language: English **Document Type:** Conference Paper (PA)

Treatment: Practical (P); Theoretical (T); Experimental (X)

Abstract: Some of the major problems in **magnetic resonance imaging** are related to static artifacts generated through interactions with the main magnetic field. The artifacts caused by biomedical implants such as orthopedic braces, prostheses or other fixation device implants distort the information contained in the image and impair the process of clinical investigation. In this paper we present preliminary results of a non-linear adjustive method for partial reduction of such type of distortion. (5 Refs)

Descriptors: biomedical **MRI**; correlation methods; medical image processing; orthopaedics; prosthetics

Identifiers: metal objects; **MR images**; nonlinearity partial reduction; distortion partial reduction; **magnetic resonance imaging**; static artifacts; main magnetic field; biomedical implants; orthopedic braces; prostheses; fixation device implants; clinical investigation; nonlinear adjustive method; cross correlation coefficients

Class Codes: **A8760I** (Medical magnetic resonance imaging and spectroscopy); **A8770E** (Patient diagnostic methods and instrumentation); **A8740** (Biomagnetism); **A8770J** (Prosthetics and other practical applications); **B7510N** (Biomedical magnetic resonance imaging and spectroscopy); **B7520E** (Prosthetics and orthotics); **B6135** (Optical, image and video signal processing); **C7330** (Biology and medical computing); **C5260B** (Computer vision and image processing techniques

Copyright 2000, IEE

34/9/5 (Item 1 from file: 155) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

MEDLINE(R)

(c) format only 2005 Dialog. All rights reserved.

15368296 PMID: 15179281

[Imaging for stereotaxic treatment of vestibular schwannomas. Error factors and corrections]

Imagerie pour le traitement stereotaxique des schwannomes vestibulaires. Facteurs d'erreur et corrections.

Wikler D; Metens T; David P; Levivier M

Centre Gamma Knife, Hopital Erasme, Universite Libre de Bruxelles. David.Wikler@ulb.ac.be

Neuro-Chirurgie (France) Jun 2004 , 50 (2-3 Pt 2) p282-8 , ISSN: 0028-3770 **Journal Code:** 0401057

Publishing Model Print

Document type: Journal Article ; English Abstract

Languages: FRENCH

Main Citation Owner: NLM

Record type: MEDLINE; Completed

Subfile: INDEX MEDICUS

BACKGROUND AND PURPOSE: Gamma Knife radiosurgery treatment of vestibular schwannomas requires high accuracy for the prescribed dose definition and delivery. The main factors contributing to the error are the anatomical distortions of imaging modalities used for treatment planning. Imaging limitations and error factors are reviewed and detailed. Multimodality rationale for the delineation of vestibular schwannomas and surrounding structures are assessed. Quality control strategies are discussed and a distortion correction technique using a radiological phantom is presented. **METHODS:** Computed tomography is considered as the reference for spatial accuracy after appropriate scanner quality control using the stereotaxic fiducials system. **Magnetic resonance imaging** pulse sequence distortions are measured with a phantom designed for 3D **non-linear** local **distortion** evidence. A distortion correction transformation is computed from the phantom images and applied to the patient images. Results are verified using the stereotaxic fiducials system. **RESULTS:** Fiducials registration errors show spatial accuracy improvement, approaching computed tomography quality, after distortion correction of **magnetic resonance images**. **CONCLUSIONS:** The multimodal imaging approach for the dose planning of vestibular schwannomas radiosurgery treatment is relevant. Quality control of spatial accuracy for imaging modalities is mandatory and realistic in clinical routine.

Descriptors: *Ear Neoplasms--surgery--SU; *Neuroma, Acoustic--surgery--SU; *Postoperative Complications--prevention and control--PC; *Radiosurgery --instrumentation--IS ; Ear Neoplasms--pathology--PA; Ear Neoplasms--radiography--RA; Humans; **Magnetic Resonance Imaging**; Neuroma, Acoustic--pathology--PA; Neuroma, Acoustic--radiography--RA; Tomography, X-Ray Computed

Record Date Created: 20040604

Record Date Completed: 20040826

34/9/6 (Item 1 from file: 8) [Links](#)

Ei Compendex(R)

(c) 2005 Elsevier Eng. Info. Inc. All rights reserved.

07457032 E.I. No: EIP05259165253

Title: Motion correction in fMRI by mapping slice-to-volume with concurrent field-inhomogeneity correction

Author: Yeo, Desmond T. B.; Fessier, Jeffery A.; Kim, Boklye

Conference Title: Medical Image Computing and Computer-Assisted Intervention, MICCAI 2004 - 7th International Conference, Proceedings

Conference Location: Saint-Malo, France **Conference Date:** 20040926-20040929

E.I. Conference No.: 64990

Source: Lecture Notes in Computer Science Medical Image Computing and Computer-Assisted Intervention, MICCAI 2004 - 7th International Conference, Proceedings v 3217 n 1 PART 2 2004.

Publication Year: 2004

ISSN: 0302-9743

Language: English

Document Type: CA; (Conference Article) **Treatment:** T; (Theoretical)

Journal Announcement: 0505W5

Abstract: Head motion is the major source of error in measuring intensity changes related to given stimuli in fMRI. The effects of head motion are image shifts and field inhomogeneity variations which cause local changes in geometric distortions. The previously developed motion correction method, mapping slice-to-volume (MSV), retrospectively remaps slices that are shifted by head motion to their spatially correct locations in an anatomical reference. Images exhibiting spatially varying geometric **distortions** require **non-linear** mapping solutions. An accurate field map can be used for the correction of such spatial distortions. However, field-map changes with head motion and, in practice, only one field-map is available typically. This work evaluates the improved motion correction capability of MSV with concurrent iterative field-corrected reconstruction using only an initial field-map. The results from simulated motion data show effective convergence and accuracy in image registration for the correction of image artifacts complicated by the motion induced field effects. copy Springer-Verlag Berlin Heidelberg 2004. 7 Refs.

Descriptors: *Magnetic resonance imaging; Mapping ; Image analysis; Data reduction; Magnetic susceptibility; Brain; Computer simulation

Identifiers: Motion correction; Map-slice-to-volume (MSV); Image artifacts; Head motion

Classification Codes:

701.2 (Magnetism, Basic Concepts & Phenomena); 461.1 (Biomedical Engineering); 723.2 (Data Processing); 723.5 (Computer Applications)

701 (Electricity & Magnetism); 461 (Bioengineering); 741 (Light, Optics & Optical Devices); 723 (Computer Software, Data Handling & Applications)

70 (ELECTRICAL ENGINEERING, GENERAL); 46 (BIOENGINEERING); 74 (LIGHT & OPTICAL TECHNOLOGY); 72 (COMPUTERS & DATA PROCESSING)

34/9/7 (Item 1 from file: 35) [Links](#)

Dissertation Abs Online

(c) 2005 ProQuest Info&Learning. All rights reserved.

01301938 ORDER NO: AADMM-74514

GEOMETRICAL DISTORTION OF MAGNETIC RESONANCE IMAGES

Author: GAUVIN, ALAIN

Degree: M.SC.

Year: 1992

Corporate Source/Institution: MCGILL UNIVERSITY (CANADA) (0781)

Source: Volume 31/03 of MASTERS ABSTRACTS. of Dissertations Abstracts International.

PAGE 1257 . 134 PAGES

Descriptors: PHYSICS, OPTICS; HEALTH SCIENCES, RADIOLOGY; ENGINEERING, BIOMEDICAL

Descriptor Codes: 0752; 0574; 0541

ISBN: 0-315-74514-2

The problem of geometrical distortion in **MR images** is addressed in the context of the applicability of stereotactic techniques. For this purpose, the distortion of phantom images is measured at various readout bandwidths and the spatial linearity is evaluated in view of the use of a stereotactic frame. The presence of a contribution to the overall **distortion** of **non- linear** magnetic gradients is shown from the data, although the distortion observed seems to be mostly attributable to the main field inhomogeneity. The specific problems of distortion of the fiducial markers due to bulk magnetic susceptibility effects is addressed. The occurrence of such effects is characterized with the help of imaging, and the role of the phenomenon on proper target localization is demonstrated. In addition, a method of bypassing the detrimental aspect of these effects is presented.

Various distortion correction approaches are discussed, and their benefits and drawbacks are evaluated. In the light of this discussion, a recently reported correction method is then presented. This method allows the calculation of an image free from geometrical and intensity distortion from the combined effect of main field inhomogeneity, susceptibility effects and chemical shift. (Abstract shortened by UMI.)

34/9/8 (Item 1 from file: 144) Links

Fulltext available through: USPTO Full Text Retrieval Options

Pascal

(c) 2005 INIST/CNRS. All rights reserved.

14624300 PASCAL No.: 00-0294719

A quantitative study of the pixel-shifting, blurring and nonlinear distortions in **MRI** images caused by the presence of metal implants

BUI F M; BOTT K; MINTCHEV M P

Department of Electrical and Computer Engineering, University of Calgary, Calgary, Alberta, T2N 1N4, Canada; GE Medical Systems, Calgary, Alberta, T2N 1N4, Canada; Department of Surgery, University of Alberta, Edmonton, Alberta, T6G 2B7, Canada

Journal: Journal of medical engineering & technology

, 2000, 24 (1

) 20-27

ISSN: 0309-1902 CODEN: JMTEDN Availability: INIST-14682

; 354000088505300040

No. of Refs.: 20 ref.

Document Type: P (Serial) ; A (Analytic)

Country of Publication: United Kingdom

Language: English

Magnetic resonance imaging has found an increasing number of medical applications in recent years due to its technical merits as well as its non-invasive nature. However, its full potential has been severely limited by magnetic susceptibility difference artefacts caused by the presence of ferromagnetic sources such as orthopedic implants, dental work or metallic needles used in neurosurgery. In this study, we propose a method to numerically quantify the distortions resulting from the magnetic susceptibility differences by investigating the phenomena from three perspectives: (1) pixel displacement, (2) blurring and (3) nonlinearity. For this purpose, phantom images obtained from a **magnetic resonance** scanner were studied. Attempts made to reconstruct an ideal image from its distorted version by appropriately compensating for the three types of distortions yielded encouraging results.

English Descriptors: **Nuclear magnetic**

resonance imaging; Interference; Implanted

material; Surgery; Metal; Quantitative analysis; Image quality;

Non linear distortion; Blurred

image; Magnetic susceptibility; Test object

Broad Descriptors: Medical imagery; Imagerie medicale; Imagerie medical

French Descriptors: Imagerie RMN; Interference; Matériau implante;

Chirurgie; Metal; Analyse quantitative; Qualité image; Distorsion non

linéaire; Image floue; Susceptibilité magnétique; Objet test

Classification Codes: 002B24A10

Copyright (c) 2000 INIST-CNRS. All rights reserved.

34/9/9 (Item 1 from file: 34) [Links](#)

Fulltext available through: [USPTO Full Text Retrieval Options](#)

SciSearch(R) Cited Ref Sci

(c) 2005 Inst for Sci Info. All rights reserved.

10609119 **Genuine Article#:** 547BQ **Number of References:** 15

A patch algorithm for fast registration of distortions

Author: Zhilkin P (REPRINT) ; Alexander ME

Corporate Source: Natl Res Council Canada,Inst Biodiagnost,435 Ellice Ave/Winnipeg/MB R3B 1Y6/Canada/ (REPRINT); Natl Res Council Canada,Inst Biodiagnost,Winnipeg/MB R3B 1Y6/Canada/

Journal: VIBRATIONAL SPECTROSCOPY , 2002 , V 28 , N1 (FEB 28) , P 67-72

ISSN: 0924-2031 **Publication date:** 20020228

Publisher: ELSEVIER SCIENCE BV , PO BOX 211, 1000 AE AMSTERDAM, NETHERLANDS

Language: English **Document Type:** ARTICLE

Geographic Location: Canada

Journal Subject Category: CHEMISTRY, ANALYTICAL; CHEMISTRY, PHYSICAL; SPECTROSCOPY

Abstract: This note describes an algorithm for registration of **non- linear distortions** in two-dimensional (2D) images. The algorithm solves an integrated for-m of linearized image matching equation over a set of patches in the image domain. The distortion field is represented as a linear combination of multinomials in the image-plane coordinates. Prior to registration, the images are convolved with band-pass filters that are implemented at a coarse-to-fine hierarchy of spatial scales (specified at input by the user); and some heuristics are given for choosing these scales. Images registered at a coarse scale are used as input approximations for deriving images registered at a subsequent finer scale of resolution. With this prescribed hierarchy, the algorithm is non-iterative it is also robust to noise, and can in principle accommodate arbitrary local or global distortions. The results indicate that non-linear registration does improve the match between images. (C) 2002 Elsevier Science B.V. All rights reserved.

Descriptors--Author Keywords: image processing ; non-linear registration ; multiscale filtering

Identifiers-- KeyWord Plus(R): IMAGE REGISTRATION; **MR-IMAGES**

Cited References:

- ALEXANDER ME, 1998, P2090, P 6 ISMRM
- ALEXANDER ME, 1999, V10, P242, INT J IMAG SYST TECH
- FRISTON KJ, 1995, V2, P165, HUMAN BRAIN MAPPING
- GOLDSTEIN SR, 1997, V16, P17, IEEE T MED IMAGING
- GUPTA N, 1997, V22, P81, INT J COMPUT VISION
- HAJNAL JV, 1995, V19, P289, J COMPUT ASSIST TOMO
- HILL DLG, 2001, V46, PR1, PHYS MED BIOL
- HORN BKP, 1981, V17, P185, ARTIF INTELL
- PRESS WH, 1992, NUMERICAL RECIPES C
- RUECKERT D, 1999, V18, P712, IEEE T MED IMAGING
- SIMONCELLI EP, 1994, P790, P ICIP 94 AUST TX
- WOODS RP, 1998, V22, P141, J COMPUT ASSIST TOMO
- ZHILKIN P, 2001, P120, P ISMRM GLASG SCOTL
- ZHILKIN P, 1999, P128, P 7 ISMRM PHIL PA
- ZHILKIN P, 2000, V18, P1143, MAGN RESON IMAGING

34/9/10 (Item 2 from file: 34) [Links](#)

Fulltext available through: [John Wiley and Sons](#) [USPTO Full Text Retrieval Options](#)

SciSearch(R) Cited Ref Sci

(c) 2005 Inst for Sci Info. All rights reserved.

07766228 **Genuine Article#:** 206FQ **Number of References:** 21

MRI geometric distortion: A simple approach to correcting the effects of non-linear gradient fields

Author: Langlois S (REPRINT) ; Desvignes M; Constans JM; Revenu M

Corporate Source: ISMRA UNIV,GREYC, 6 BD MARECHAL JUIN/F-14050 CAEN//FRANCE/ (REPRINT)

Journal: JMRI-JOURNAL OF MAGNETIC RESONANCE IMAGING , 1999 , V 9 , N6 (JUN) , P 821-831

ISSN: 1053-1807 **Publication date:** 19990600

Publisher: JOHN WILEY & SONS INC , 605 THIRD AVE, NEW YORK, NY 10158-0012

Language: English **Document Type:** ARTICLE

Geographic Location: FRANCE

Subfile: CC CLIN--Current Contents, Clinical Medicine

Journal Subject Category: RADIOLOGY, NUCLEAR MEDICINE & MEDICAL IMAGING

Abstract: We present a method to correct intensity variations and voxel shifts caused by non-linear gradient fields in **magnetic resonance images**. The principal sources of distortion are briefly discussed, as well as the methods of correction currently in use. The implication of the gradient held non-linearities on the signal equations are described in a detailed way for the case of two- and three-dimensional Fourier imaging. A model of these non-linearities, derived from the geometry of the gradient coils, is proposed and then applied in post-processing to correct any images regardless of the acquisition sequence. Initial position errors, as large as 4 mm (i.e., four voxels of 1 x 1 x 1.4 mm(3)) before correction, are reduced to less than the voxel sizes after correction. (C) 1999 Wiley-Liss, Inc.

Descriptors--Author Keywords: **magnetic resonance imaging ; image distortion correction ; non-linear gradient fields**

Identifiers-- KeyWord Plus(R): **MAGNETIC-RESONANCE IMAGES; ACCURACY**

Cited References:

- BANDETTINI PA, 1995, V6, P133, INT J IMAG SYST TECH
- DURAND E, 1966, V1, ELECTROMAGNETISME
- GUINET J, 1992, INTRO IRM
- HAACKE EM, 1990, V155, P951, AM J ROENTGENOL
- HINSHAW WS, 1983, V71, P338, P IEEE
- HUGHES DG, 1992, V25, P167, MAGNET RESON MED
- KORACH G, 1992, MANUEL TECHNIQUES IM
- MAURER CR, 1996, V20, P666, J COMPUT ASSIST TOMO
- MICHIELS J, 1994, V12, P749, MAGN RESON IMAGING
- MOERLAND MA, 1995, V40, P1651, PHYS MED BIOL
- POSSE S, 1992, V25, P12, MAGNET RESON MED
- PRESS WH, 1992, NUMERICAL RECIPES C
- SCHAD L, 1987, V11, P499, J COMPUT ASSIST TOMO
- SCHENCK JF, 1996, V23, P815, MED PHYS
- SUMANaweera T, 1994, V31, P40, MAGNET RESON MED
- SUMANaweera TS, 1994, V35, P1, NEUROSURGERY
- SUN L, 1994, V31, P359, MAGNET RESON MED
- VRE RM, 1990, P77, MAGNETIC RESONANCE S
- WALTON L, 1996, V38, P170, NEUROSURGERY
- WELLS WM, 1996, V15, P429, IEEE T MED IMAGING

ZHU XX, 1995, V4, P217, SOLID STATE NUCL MAG

34/9/11 (Item 1 from file: 89) [Links](#)

GeoRef

(c) 2005 American Geological Institute. All rights reserved.

01631660 Georef No.: 89-54004

Title: Enhancing borehole image data on a high-resolution PC

Author: Wong, S. A.; Startzman, R. A.; Kuo, T. B.

Corporate Source: Tex. A&M Univ., College Station, TX, United States

Monograph Title: Proceedings; Petroleum computer conference

Author: Anonymous

Conference Title: Petroleum computer conference

Conference Location: San Antonio, TX, United States,

Conference Date: June 25-28, 1989

Publisher: Soc. Pet. Eng., Richardson, TX, United States

p. 37-48

Date: 1989

Country of Publication: United States

Refs.: 22

Document Type: Book; Conference document

Bibliographic Level: Analytic

Illustrations: illus. incl. 3 tables

Language: English

Descriptors: boreholes; computers; convolution; data processing; downhole methods; engineering geology; filters; geophysical methods; high-resolution methods; imagery; methods; microcomputers; **non-linear distortion**; **nuclear magnetic resonance**; pattern recognition; personal computers; petroleum engineering; reservoir properties; seismic methods; spectroscopy; well-logging

Section Headings: 30 (Engineering geology); 20 (Applied geophysics);

Georef Update: 1989

GeoRef, Copyright 1995, American Geological Institute.

34/9/12 (Item 1 from file: 65) [Links](#)

Inside Conferences

(c) 2005 BLDSC all rts. reserv. All rights reserved.

03154131 **Inside Conference Item ID:** CN033417287

Method For Partial Reduction of the Non-linear and Distortion in MR Images Caused by the Presence of Metal Objects

Bott, K.; Mintchev, M.; Smith, M.; Domoustchiev, K.

Conference: Electrical and computer engineering; Engineering solutions for the next millennium - Canadian conference; 12th

CANADIAN CONFERENCE ON ELECTRICAL AND COMPUTER ENGINEERING , 1999; VOL 2 P: 851-854
IEEE, 1999

ISSN: 0840-7789 **ISBN:** 0780355792

Language: English **Document Type:** Conference Papers and programme

Editor: Meng, M.

Sponsor: IEEE Canada

Location: Edmonton, Canada

Date: May 1999 (199905) (199905)

British Library Item Location: 3019.445500

Note:

Also known as CCECE '99. IEEE cat no 99TH8411

Descriptors: electrical engineering; computer engineering; engineering; CCECE; IEEE

34/9/13 (Item 1 from file: 347) [Links](#)

JAPIO

(c) 2005 JPO & JAPIO. All rights reserved.

05986411 **Image available**

METHOD OF MEASURING NONLINEAR DISTORTION

Pub. No.: 10-269511 [JP 10269511 A]

Published: October 09, 1998 (19981009)

Inventor: HONDA NAOKI

KOMAGINE TAKASHI

OUCHI KAZUHIRO

Applicant: AKITA PREF GOV [000000] (A Japanese Government or Municipal Agency), JP (Japan)

Application No.: 09-075652 [JP 9775652]

Filed: March 27, 1997 (19970327)

International Class: [6] G11B-005/09; G11B-005/09; **G01R-033/12**

JAPIO Class: 42.5 (ELECTRONICS -- Equipment); 46.2 (INSTRUMENTATION -- Testing)

ABSTRACT

PROBLEM TO BE SOLVED: To provide such a method of measuring nonlinear distortion of a digital record, as a phase distortion and an amplitude distortion can be measured separately with precision, a direction of a phase distortion is identified, and this measured phase distortion value can be used as a record pre-correction quantity.

SOLUTION: When a nonlinear distortion is measured at magnetic recording, this method is to provide a **non-linear distortion** measuring method by which phase distortion (Δ) and amplitude distortion (r) are separated and also with precision, for example, in a phase distortion measurement, phase distortion quantity in a desired recording density is obtained with precision and also a direction of the distortion is obtained by directly precorrecting dibit part of a recording pattern for recording. Moreover, from this quantity, the amplitude distortion quantity (r) in the recording density can be calculated with precision. It is possible to use the pre-correction quantity used for the measurement then as an optimal pre-correction quantity for a control at a time of recording operation.

34/9/14 (Item 2 from file: 347) [Links](#)

JAPIO

(c) 2005 JPO & JAPIO. All rights reserved.

03046345 **Image available**

PHASE DISTORTION CORRECTING METHOD IN MAGNETIC RESONANCE IMAGING DEVICE

Pub. No.: 02-021845 [JP 2021845 A]

Published: January 24, 1990 (19900124)

Inventor: MAEDA AKIRA

KASAMA TAKASHI

YOKOYAMA TETSUO

NISHIMURA HIROSHI

Applicant: HITACHI LTD [000510] (A Japanese Company or Corporation), JP (Japan)

HITACHI MICRO COMPUT ENG LTD [470864] (A Japanese Company or Corporation), JP (Japan)

HITACHI MEDICAL CORP [420143] (A Japanese Company or Corporation), JP (Japan)

Application No.: 63-170928 [JP 88170928]

Filed: July 11, 1988 (19880711)

International Class: [5] A61B-005/05; G01R-033/28

JAPIO Class: 28.2 (SANITATION -- Medical)

Journal: Section: C, Section No. 706, Vol. 14, No. 171, Pg. 12, April 03, 1990 (19900403)

ABSTRACT

PURPOSE: To lessen a large **non-linear phase distortion** of the whole complex image by setting an partial region (block) of an arbitrary form on a reconstructed complex image, and dividing the complex image into blocks.

CONSTITUTION: In a rotary recovery method which is one of imaging methods in the **magnetic resonance imaging**, image $C'(\text{sub } 1k) = (C(\text{sub } 1k))^{(\text{sup } 2)}$ where reconstructed image containing phase distortion is squared is calculated. An absolute value for $C'(\text{sub } 1k)$ is taken and the maximum value (CMAX) is calculated. $C'(\text{sub } 1k)$ is divided into blocks $C'(\text{sub } mn)$. The size of the block is decided so that the change of the phase distortion is kept within the range of plus or minus π . if possible. For example, in case of an original image of 256X256 picture elements, if it is divided into 4X4 blocks, the block size is 64X64. The number of elements to satisfy CMAX.TH (predetermined threshold value) in $C'(\text{sub } mn)$ and the predetermined number of picture elements are compared and decided which is larger, and when $M > (\text{the number of picture elements satisfying CMAX.TH})$, the block $C'(\text{sub } mn)$ is Fourier transformed to obtain $B'(\text{sub } xy)$. The absolute value of $B'(\text{sub } xy)$ is taken to calculate the coordinate (x', y') of the maximum value. Further, the phase angle A in (x', y') is calculated.

34/9/15 (Item 3 from file: 347) [Links](#)

JAPIO

(c) 2005 JPO & JAPIO. All rights reserved.

01227671 ****Image available****

MEASUREMENT OF DC MAGNETIC FIELD

Pub. No.: 58-165071 [JP 58165071 A]

Published: September 30, 1983 (19830930)

Inventor: YOSHINO TOSHIHIKO

ITO KATSUJI

Applicant: FURUKAWA ELECTRIC CO LTD THE [000529] (A Japanese Company or Corporation), JP (Japan)
YOSHINO TOSHIHIKO [000000] (An Individual), JP (Japan)

Application No.: 57-046376 [JP 8246376]

Filed: March 25, 1982 (19820325)

International Class: [3] G01R-033/032

JAPIO Class: 46.1 (INSTRUMENTATION -- Measurement)

JAPIO Keyword: R002 (LASERS); R012 (OPTICAL FIBERS)

Journal: Section: P, Section No. 246, Vol. 08, No. 1, Pg. 6, January 06, 1984 (19840106)

ABSTRACT

PURPOSE: To measure a DC magnetic field at a high sensitivity and a high accuracy by detecting **distortion** of a **non- linear** magnetostrictive element with a Fabry-Perot interferometer made of an optical fiber wound around the element according to the DC magnetic field.

CONSTITUTION: A non-linear magnetostrictive element 10 to which an AC bias magnetic field is applied through a bias coil 14 is distorted greatly and deformed according to a DC magnetic field from an air core coil 26. This distortion is detected at a high sensitivity and a high resolution with a Fabry-Perot interferometer 16 made of an optical fiber wound around the element 10 and the DC magnetic field measured with a light receiving photomultiplier 20 and an oscilloscope 22 or a counter 24 or the like at a high sensitivity and high accuracy